



More information on the website
radwag.com/en/info,w1,9XE

100 μ l ÷ 1000 μ l - RW8-106-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	100 – 1000 μ l
Systematic error	2.0% / 0.6%
Random error	0.7% / 0.2%
Tips	1000 μ l



Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration