



More information on the website
radwag.com/us/info,w1,9XE

100 μ l \div 1000 μ l - RW8-106-20-9 Automatic Pipettes

PA-203-0006



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Volume	100 – 1000 μ l
Systematic error	2,0% / 0,6%
Random error	0,7% / 0,2%
Tips	1000 μ l



Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration