



More information on the website
radwag.com/en/info,w1,A7T

2 μ l \div 20 μ l - RW8-102-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters	
Volume	2 μ l \div 20 μ l
Systematic error	3.0% / 0.9%
Random error	2.0% / 0.4%
Tips	300 μ l



Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration
Microbalance for Pipette Calibration MYA 5Y.P

Automatic Balances for. Pipettes Calibration
Pipette Tips