



More information on the website  
[radwag.com/en/info,w1,IRH](http://radwag.com/en/info,w1,IRH)

# 20 $\mu$ l ÷ 200 $\mu$ l - RW8-105-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

Metrological parameters	
Volume	20 – 200 $\mu$ l
Systematic error	2.0% / 0.6%
Random error	0.8% / 0.15%
Tips	300 $\mu$ l



## Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration