



More information on the website
radwag.com/en/info,w1,IRH

20 μ l ÷ 200 μ l - RW8-105-20-9 Automatic Pipettes

PA-203-0005



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters

Volume	20 – 200 μ l
Systematic error	2.0% / 0.6%
Random error	0.8% / 0.15%
Tips	300 μ l



Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration