



More information on the website
radwag.com/us/info,w1,IRH

20 μ l ÷ 200 μ l - RW8-105-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Volume	20 – 200 μ l
Systematic error	2,0% / 0,6%
Random error	0,8% / 0,15%
Tips	300 μ l



Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration