



More information on the website  
[radwag.com/en/info,w1,QLH](http://radwag.com/en/info,w1,QLH)

# 5 $\mu$ l ÷ 50 $\mu$ l - RW8-103-20-9 Automatic Pipettes

PA-203-0003



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

### Metrological parameters

Volume	5 – 50 $\mu$ l
Systematic error	2.0% / 0.6%
Random error	2.0% / 0.3%
Tips	300 $\mu$ l



## Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration  
MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration  
Pipette Tips