



More information on the website  
[radwag.com/us/info,w1,QLH](https://radwag.com/us/info,w1,QLH)

# 5 $\mu$ l ÷ 50 $\mu$ l - RW8-103-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

Volume	5 $\mu$ l ÷ 50 $\mu$ l
Systematic error	2,0% / 0,6%
Random error	2,0% / 0,3%
Tips	300 $\mu$ l



## Compatible with

XA 5Y.M.A.P Microbalance  
Microbalance for Pipette Calibration MYA 5Y.P

Automatic Balances for. Pipettes Calibration  
Pipette Tips