



More information on the website  
[radwag.com/en/info,w1,UAW](http://radwag.com/en/info,w1,UAW)

# 0.5 $\mu$ l ÷ 10 $\mu$ l - RW8-101-20-9 Automatic Pipettes

PA-203-0001



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

### Metrological parameters

Volume	0.5 – 10 $\mu$ l
Systematic error	2.5% / 1.0%
Random error	1.5% / 0.8%
Tips	10 $\mu$ l



## Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Pipette Tips  
 Automatic Balances for. Pipettes Calibration