



More information on the website
radwag.com/en/info,w1,UAW

0.5 μ l \div 10 μ l - RW8-101-20-9 Automatic Pipettes

PA-203-0001



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Metrological parameters

Volume	0.5 – 10 μ l
Systematic error	2.5% / 1.0%
Random error	1.5% / 0.8%
Tips	10 μ l



Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
 MYA 5Y.P Microbalance for pipettes calibration

Pipette Tips
 Automatic Balances for. Pipettes Calibration