



More information on the website
radwag.com/us/info,w1,UAW

0,5 µl ÷ 10 µl - RW8-101-20-9 Automatic Pipettes

PA-203-0001



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

| | |
|------------------|-------------|
| Volume | 0,5 – 10 µl |
| Systematic error | 2,5% / 1,0% |
| Random error | 1,5% / 0,8% |
| Tips | 10 µl |



Compatible with (Additional Fee)

XA 5Y.M.A.P Microbalance for pipettes calibration
 MYA 5Y.P Microbalance for pipettes calibration

Pipette Tips
 Automatic Balances for. Pipettes Calibration