



More information on the website  
[radwag.com/us/info,w1,VML](http://radwag.com/us/info,w1,VML)

# 500 $\mu$ l ÷ 5000 $\mu$ l - RW8-107-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

## Datasheet

Volume	500 – 5000 $\mu$ l
Systematic error	2,0% / 0,5%
Random error	0,6% / 0,15%
Tips	5000 $\mu$ l



## Compatible with

XA 5Y.M.A.P Microbalance for pipettes calibration  
 MYA 5Y.P Microbalance for pipettes calibration

Automatic Balances for. Pipettes Calibration