



More information on the website
radwag.com/us/info,w1,VML

500 μ l ÷ 5000 μ l - RW8-107-20-9 Automatic Pipettes



The drawings, photos and graphics used are for illustrative purposes only.

Datasheet

Volume	500 μ l ÷ 5000 μ l
Systematic error	2,0% / 0,5%
Random error	0,6% / 0,15%
Tips	5000 μ l



Compatible with

XA 5Y.M.A.P Microbalance
Microbalance for Pipette Calibration MYA 5Y.P

Automatic Balances for. Pipettes Calibration